

Carrie McCrillis
Oilfield Environmental & Compliance
307 Roemer Way
Suite 300
Santa Maria, CA 93454

WP-195



Final Report

WatR™ Pollution Proficiency Testing

WatR™ Pollution Study

Open Date: 04/11/11

Close Date: 05/26/11

Report Issued Date: 06/16/11



June 17, 2011

Carrie McCrillis
Oilfield Environmental & Compliance
307 Roemer Way
Suite 300
Santa Maria, CA 93454

Enclosed is your final report for ERA's WP-195 WatR™ Pollution Proficiency Testing (PT) study. Your final report includes an evaluation of all results submitted by your laboratory to ERA.

Data Evaluation Protocols: All analytes in ERA's WP-195 WatR™ Pollution Proficiency Testing study have been evaluated using the following tiered approach. If the analyte is listed in the most current National Environmental Laboratory Accreditation Conference (NELAC) PT Field of Testing tables, the evaluation was completed by comparing the reported result to the acceptance limits generated using the criteria contained in the NELAC FoPT tables. If the analyte is not included in the NELAC FoPT tables, the reported result has been evaluated using the procedures outlined in ERA's Standard Operating Procedure for the Generation of Performance Acceptance Limits (SOP 0260).

Corrective Action Help: As part of your accreditation(s), you may be required to identify the root cause of any "Not Acceptable" results, implement the necessary corrective actions, and then satisfy your PT requirements by participating in a Supplemental (QuiK™ Response) or future ERA PT study. ERA's technical staff is available to help your laboratory resolve any technical issues that may be impairing your PT performance and possibly affecting your routine data quality. Our laboratory and technical staff have well over three hundred years of collective experience in performing the full range of environmental analyses. As part of our technical support, ERA offers QC samples that can be helpful in helping you work through your technical issues.

Thank you for your participation in ERA's WP-195 WatR™ Pollution Proficiency Testing study. If you have any questions, please contact the proficiency testing department, or myself, at 1-800-372-0122.

Sincerely,

Jay R. McBurney
Quality Program Manager

attachments
jrm



A Waters Company

Report Recipient	Contact/Phone Number	Reporting Type
California	Fred Choske / 510-620-3175	All Analytes

WP-195 Definitions & Study Discussion

Study Dates: 04/11/11 - 05/26/11

Report Issued: 06/16/11

WP Study Definitions

The Reported Value is the value that the laboratory reported to ERA.

The ERA Assigned Values are compliant with the most current USEPA/NELAC FoPT tables. The assigned values are directly traceable to the commercially prepared starting materials used to manufacture the PT standards. A parameter not added to the standard is given an Assigned Value of "0" per the guidelines contained in the USEPA's Criteria Document and NELAC standards.

The Acceptance Limits are established per the criteria contained in the most current USEPA/NELAC FoPT tables, or ERA's SOP for the Generation of Performance Acceptance Limits™ as applicable.

The Performance Evaluation:

- Acceptable = Reported Value falls within the Acceptance Limits.
- Not Acceptable = Reported Value falls outside the Acceptance Limits.
- No Evaluation = Reported Value cannot be evaluated.
- Not Reported = No Value reported.

The Method Description is the method the laboratory reported to ERA.

WP Study Discussion

ERA's WP-195 WatR™ Pollution Proficiency Testing study has been reviewed by ERA senior management and certified compliant with the requirements of the USEPA's National Standards for Water Proficiency Testing Studies Criteria Document (December 1998), and the criteria contained in the most current NELAC FoPT tables.

ERA's WP-195 WatR™ Pollution study standards were examined for any anomalies. A full review of all homogeneity, stability and accuracy verification data was completed. All analytical verification data for all analytes met the acceptance criteria contained in the USEPA's National Criteria Document for Water Proficiency Testing Studies, December 1998, and the criteria contained in the most current NELAC FoPT tables.

The data submitted by participating laboratories was also examined for study anomalies. There were no anomalies observed during the statistical review of the data.

ERA's WP-195 WatR™ Pollution study reports shall not be reproduced except in their entirety and not without the permission of the participating laboratories. The report must not be used by the participating laboratories to claim product endorsement by any agency of the U. S. government.

The data contained herein are confidential and intended for your use only.

If you have any questions or concerns regarding your assessment in ERA's WatR™ Pollution Proficiency Testing program, please contact Jay McBurney, Quality Program Manager, or the proficiency testing department at 1-800-372-0122.



Study: **WP-195**

ERA Customer Number: **O460853**

Laboratory Name: **Oilfield Environmental &
Compliance**

Inorganic Results





A Waters Company

WP-195 Final Complete Report

Carrie McCrillis
Quality Assurance Officer
Oilfield Environmental & Compliance
307 Roemer Way
Suite 300
Santa Maria, CA 93454
(805) 922-4772

EPA ID: CA01480
ERA Customer Number: O460853
Report Issued: 06/16/11
Study Dates: 04/11/11 - 05/26/11

Anal. No.	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description
WP Complex Nutrients (cat# 579)							
0034	Total Kjeldahl Nitrogen	mg/L	12.4	13.6	9.02 - 17.6	Acceptable	SM4500NH3 D
0035	Total phosphorus as P	mg/L		2.18	1.75 - 2.66	Not Reported	



Study: **WP-195**

ERA Customer Number: **O460853**

Laboratory Name: **Oilfield Environmental &
Compliance**

Organic Results





A Waters Company

WP-195 Final Complete Report

Carrie McCrillis
Quality Assurance Officer
Oilfield Environmental & Compliance
307 Roemer Way
Suite 300
Santa Maria, CA 93454
(805) 922-4772

EPA ID:
ERA Customer Number:
Report Issued:
Study Dates:

CA01480
O460853
06/16/11
04/11/11 - 05/26/11

Anal. No.	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description
-----------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------

WP Volatiles (cat# 830)

4315	Acetone	µg/L		48.2	9.34 - 78.0	Not Reported	
4320	Acetonitrile	µg/L		0.00		Not Reported	
4325	Acrolein	µg/L		0.00		Not Reported	
4340	Acrylonitrile	µg/L		0.00		Not Reported	
0065	Benzene	µg/L		18.9	12.8 - 25.0	Not Reported	
0060	Bromodichloromethane	µg/L		70.0	49.6 - 94.5	Not Reported	
0062	Bromoform	µg/L		0.00		Not Reported	
4950	Bromomethane	µg/L		32.0	12.8 - 51.2	Not Reported	
4410	2-Butanone (MEK)	µg/L		0.00		Not Reported	
5000	tert-Butyl methyl ether (MTBE)	µg/L		0.00		Not Reported	
4450	Carbon disulfide	µg/L		0.00		Not Reported	
0058	Carbon tetrachloride	µg/L		0.00		Not Reported	
0064	Chlorobenzene	µg/L		63.4	45.7 - 79.4	Not Reported	
0061	Chlorodibromomethane	µg/L		85.2	58.4 - 113	Not Reported	
4485	Chloroethane	µg/L		0.00		Not Reported	
4500	2-Chloroethylvinylether	µg/L		0.00		Not Reported	
0055	Chloroform	µg/L		45.9	31.7 - 59.5	Not Reported	
4960	Chloromethane	µg/L		0.00		Not Reported	
4570	1,2-Dibromo-3-chloropropane (DBCP)	µg/L		0.00		Not Reported	
4585	1,2-Dibromoethane (EDB)	µg/L		0.00		Not Reported	
4595	Dibromomethane	µg/L		0.00		Not Reported	
0094	1,2-Dichlorobenzene	µg/L		70.3	48.8 - 91.2	Not Reported	
0096	1,3-Dichlorobenzene	µg/L		52.2	35.2 - 66.6	Not Reported	
0095	1,4-Dichlorobenzene	µg/L		0.00		Not Reported	
4625	Dichlorodifluoromethane (Freon 12)	µg/L		0.00		Not Reported	
4630	1,1-Dichloroethane	µg/L		17.0	11.1 - 23.1	Not Reported	
0054	1,2-Dichloroethane	µg/L		17.5	12.1 - 24.0	Not Reported	
4640	1,1-Dichloroethylene	µg/L		34.2	17.9 - 49.8	Not Reported	
4645	cis-1,2-Dichloroethylene	µg/L		28.8	20.0 - 38.5	Not Reported	
4700	trans-1,2-Dichloroethylene	µg/L		0.00		Not Reported	
4655	1,2-Dichloropropane	µg/L		0.00		Not Reported	
4680	cis-1,3-Dichloropropylene	µg/L		18.7	13.1 - 24.3	Not Reported	





A Waters Company

WP-195 Final Complete Report

Carrie McCrillis
Quality Assurance Officer
Oilfield Environmental & Compliance
307 Roemer Way
Suite 300
Santa Maria, CA 93454
(805) 922-4772

EPA ID:
ERA Customer Number:
Report Issued:
Study Dates:

CA01480
O460853
06/16/11
04/11/11 - 05/26/11

Anal. No.	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description
-----------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------

WP Volatiles (cat# 830) (Continued)

4685	trans-1,3-Dichloropropylene	µg/L	65.8	57.2	36.7 - 77.3	Acceptable	EPA 8260B
0066	Ethylbenzene	µg/L		0.00		Not Reported	
4835	Hexachlorobutadiene	µg/L		0.00		Not Reported	
4860	2-Hexanone	µg/L		78.3	39.2 - 114	Not Reported	
0063	Methylene chloride	µg/L		52.7	32.2 - 73.7	Not Reported	
4995	4-Methyl-2-pentanone (MIBK)	µg/L		67.5	30.2 - 102	Not Reported	
5005	Naphthalene	µg/L		0.00		Not Reported	
5100	Styrene	µg/L		76.7	48.9 - 105	Not Reported	
5105	1,1,1,2-Tetrachloroethane	µg/L		0.00		Not Reported	
5110	1,1,2,2-Tetrachloroethane	µg/L		25.7	13.5 - 39.9	Not Reported	
0059	Tetrachloroethylene	µg/L		62.2	34.8 - 81.3	Not Reported	
0067	Toluene	µg/L		0.00		Not Reported	
5155	1,2,4-Trichlorobenzene	µg/L		67.0	13.8 - 82.0	Not Reported	
0056	1,1,1-Trichloroethane	µg/L		28.2	17.7 - 37.7	Not Reported	
5165	1,1,2-Trichloroethane	µg/L		0.00		Not Reported	
0057	Trichloroethylene	µg/L		63.2	40.2 - 82.5	Not Reported	
5175	Trichlorofluoromethane	µg/L		0.00		Not Reported	
5180	1,2,3-Trichloropropane (TCP)	µg/L		0.00		Not Reported	
5225	Vinyl acetate	µg/L		0.00		Not Reported	
5235	Vinyl chloride	µg/L		28.4	11.4 - 45.4	Not Reported	
5260	Xylenes, total	µg/L		150	86.0 - 201	Not Reported	





A Waters Company

WP-195 Final Complete Report

Carrie McCrillis
Quality Assurance Officer
Oilfield Environmental & Compliance
307 Roemer Way
Suite 300
Santa Maria, CA 93454
(805) 922-4772

EPA ID:
ERA Customer Number:
Report Issued:
Study Dates:

CA01480
O460853
06/16/11
04/11/11 - 05/26/11

Anal. No.	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description
-----------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------

WP Volatiles (cat# 830)

4315	Acetone	µg/L		48.2	9.34 - 78.0	Not Reported	
4320	Acetonitrile	µg/L		0.00		Not Reported	
4325	Acrolein	µg/L		0.00		Not Reported	
4340	Acrylonitrile	µg/L		0.00		Not Reported	
0065	Benzene	µg/L		18.9	12.8 - 25.0	Not Reported	
0060	Bromodichloromethane	µg/L		70.0	49.6 - 94.5	Not Reported	
0062	Bromoform	µg/L		0.00		Not Reported	
4950	Bromomethane	µg/L		32.0	12.8 - 51.2	Not Reported	
4410	2-Butanone (MEK)	µg/L		0.00		Not Reported	
5000	tert-Butyl methyl ether (MTBE)	µg/L		0.00		Not Reported	
4450	Carbon disulfide	µg/L		0.00		Not Reported	
0058	Carbon tetrachloride	µg/L		0.00		Not Reported	
0064	Chlorobenzene	µg/L		63.4	45.7 - 79.4	Not Reported	
0061	Chlorodibromomethane	µg/L		85.2	58.4 - 113	Not Reported	
4485	Chloroethane	µg/L		0.00		Not Reported	
4500	2-Chloroethylvinylether	µg/L		0.00		Not Reported	
0055	Chloroform	µg/L		45.9	31.7 - 59.5	Not Reported	
4960	Chloromethane	µg/L		0.00		Not Reported	
4570	1,2-Dibromo-3-chloropropane (DBCP)	µg/L		0.00		Not Reported	
4585	1,2-Dibromoethane (EDB)	µg/L		0.00		Not Reported	
4595	Dibromomethane	µg/L		0.00		Not Reported	
0094	1,2-Dichlorobenzene	µg/L		70.3	48.8 - 91.2	Not Reported	
0096	1,3-Dichlorobenzene	µg/L		52.2	35.2 - 66.6	Not Reported	
0095	1,4-Dichlorobenzene	µg/L		0.00		Not Reported	
4625	Dichlorodifluoromethane (Freon 12)	µg/L		0.00		Not Reported	
4630	1,1-Dichloroethane	µg/L		17.0	11.1 - 23.1	Not Reported	
0054	1,2-Dichloroethane	µg/L		17.5	12.1 - 24.0	Not Reported	
4640	1,1-Dichloroethylene	µg/L		34.2	17.9 - 49.8	Not Reported	
4645	cis-1,2-Dichloroethylene	µg/L		28.8	20.0 - 38.5	Not Reported	
4700	trans-1,2-Dichloroethylene	µg/L		0.00		Not Reported	
4655	1,2-Dichloropropane	µg/L		0.00		Not Reported	
4680	cis-1,3-Dichloropropylene	µg/L		18.7	13.1 - 24.3	Not Reported	





A Waters Company

WP-195 Final Complete Report

Carrie McCrillis
Quality Assurance Officer
Oilfield Environmental & Compliance
307 Roemer Way
Suite 300
Santa Maria, CA 93454
(805) 922-4772

EPA ID:
ERA Customer Number:
Report Issued:
Study Dates:

CA01480
O460853
06/16/11
04/11/11 - 05/26/11

Anal. No.	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description
-----------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------

WP Volatiles (cat# 830) (Continued)

4685	trans-1,3-Dichloropropylene	µg/L	65.8	57.2	36.7 - 77.3	Acceptable	EPA 624
0066	Ethylbenzene	µg/L		0.00		Not Reported	
4835	Hexachlorobutadiene	µg/L		0.00		Not Reported	
4860	2-Hexanone	µg/L		78.3	39.2 - 114	Not Reported	
0063	Methylene chloride	µg/L		52.7	32.2 - 73.7	Not Reported	
4995	4-Methyl-2-pentanone (MIBK)	µg/L		67.5	30.2 - 102	Not Reported	
5005	Naphthalene	µg/L		0.00		Not Reported	
5100	Styrene	µg/L		76.7	48.9 - 105	Not Reported	
5105	1,1,1,2-Tetrachloroethane	µg/L		0.00		Not Reported	
5110	1,1,2,2-Tetrachloroethane	µg/L		25.7	13.5 - 39.9	Not Reported	
0059	Tetrachloroethylene	µg/L		62.2	34.8 - 81.3	Not Reported	
0067	Toluene	µg/L		0.00		Not Reported	
5155	1,2,4-Trichlorobenzene	µg/L		67.0	13.8 - 82.0	Not Reported	
0056	1,1,1-Trichloroethane	µg/L		28.2	17.7 - 37.7	Not Reported	
5165	1,1,2-Trichloroethane	µg/L		0.00		Not Reported	
0057	Trichloroethylene	µg/L		63.2	40.2 - 82.5	Not Reported	
5175	Trichlorofluoromethane	µg/L		0.00		Not Reported	
5180	1,2,3-Trichloropropane (TCP)	µg/L		0.00		Not Reported	
5225	Vinyl acetate	µg/L		0.00		Not Reported	
5235	Vinyl chloride	µg/L		28.4	11.4 - 45.4	Not Reported	
5260	Xylenes, total	µg/L		150	86.0 - 201	Not Reported	





A Waters Company

WP-195 Final Complete Report

Carrie McCrillis
Quality Assurance Officer
Oilfield Environmental & Compliance
307 Roemer Way
Suite 300
Santa Maria, CA 93454
(805) 922-4772

EPA ID:
ERA Customer Number:
Report Issued:
Study Dates:

CA01480
O460853
06/16/11
04/11/11 - 05/26/11

Anal. No.	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description
WP Volatiles (cat# 830)							
4315	Acetone	µg/L		48.2	9.34 - 78.0	Not Reported	
4320	Acetonitrile	µg/L		0.00		Not Reported	
4325	Acrolein	µg/L		0.00		Not Reported	
4340	Acrylonitrile	µg/L		0.00		Not Reported	
0065	Benzene	µg/L		18.9	12.8 - 25.0	Not Reported	
0060	Bromodichloromethane	µg/L		70.0	49.6 - 94.5	Not Reported	
0062	Bromoform	µg/L		0.00		Not Reported	
4950	Bromomethane	µg/L		32.0	12.8 - 51.2	Not Reported	
4410	2-Butanone (MEK)	µg/L		0.00		Not Reported	
5000	tert-Butyl methyl ether (MTBE)	µg/L	0.00	0.00		Acceptable	LUFT GCMS
4450	Carbon disulfide	µg/L		0.00		Not Reported	
0058	Carbon tetrachloride	µg/L		0.00		Not Reported	
0064	Chlorobenzene	µg/L		63.4	45.7 - 79.4	Not Reported	
0061	Chlorodibromomethane	µg/L		85.2	58.4 - 113	Not Reported	
4485	Chloroethane	µg/L		0.00		Not Reported	
4500	2-Chloroethylvinylether	µg/L		0.00		Not Reported	
0055	Chloroform	µg/L		45.9	31.7 - 59.5	Not Reported	
4960	Chloromethane	µg/L		0.00		Not Reported	
4570	1,2-Dibromo-3-chloropropane (DBCP)	µg/L		0.00		Not Reported	
4585	1,2-Dibromoethane (EDB)	µg/L		0.00		Not Reported	
4595	Dibromomethane	µg/L		0.00		Not Reported	
0094	1,2-Dichlorobenzene	µg/L		70.3	48.8 - 91.2	Not Reported	
0096	1,3-Dichlorobenzene	µg/L		52.2	35.2 - 66.6	Not Reported	
0095	1,4-Dichlorobenzene	µg/L		0.00		Not Reported	
4625	Dichlorodifluoromethane (Freon 12)	µg/L		0.00		Not Reported	
4630	1,1-Dichloroethane	µg/L		17.0	11.1 - 23.1	Not Reported	
0054	1,2-Dichloroethane	µg/L		17.5	12.1 - 24.0	Not Reported	
4640	1,1-Dichloroethylene	µg/L		34.2	17.9 - 49.8	Not Reported	
4645	cis-1,2-Dichloroethylene	µg/L		28.8	20.0 - 38.5	Not Reported	
4700	trans-1,2-Dichloroethylene	µg/L		0.00		Not Reported	
4655	1,2-Dichloropropane	µg/L		0.00		Not Reported	
4680	cis-1,3-Dichloropropylene	µg/L		18.7	13.1 - 24.3	Not Reported	





A Waters Company

WP-195 Final Complete Report

Carrie McCrillis
Quality Assurance Officer
Oilfield Environmental & Compliance
307 Roemer Way
Suite 300
Santa Maria, CA 93454
(805) 922-4772

EPA ID:
ERA Customer Number:
Report Issued:
Study Dates:

CA01480
O460853
06/16/11
04/11/11 - 05/26/11

Anal. No.	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description
-----------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------

WP Volatiles (cat# 830) (Continued)

4685	trans-1,3-Dichloropropylene	µg/L		57.2	36.7 - 77.3	Not Reported	
0066	Ethylbenzene	µg/L		0.00		Not Reported	
4835	Hexachlorobutadiene	µg/L		0.00		Not Reported	
4860	2-Hexanone	µg/L		78.3	39.2 - 114	Not Reported	
0063	Methylene chloride	µg/L		52.7	32.2 - 73.7	Not Reported	
4995	4-Methyl-2-pentanone (MIBK)	µg/L		67.5	30.2 - 102	Not Reported	
5005	Naphthalene	µg/L		0.00		Not Reported	
5100	Styrene	µg/L		76.7	48.9 - 105	Not Reported	
5105	1,1,1,2-Tetrachloroethane	µg/L		0.00		Not Reported	
5110	1,1,2,2-Tetrachloroethane	µg/L		25.7	13.5 - 39.9	Not Reported	
0059	Tetrachloroethylene	µg/L		62.2	34.8 - 81.3	Not Reported	
0067	Toluene	µg/L		0.00		Not Reported	
5155	1,2,4-Trichlorobenzene	µg/L		67.0	13.8 - 82.0	Not Reported	
0056	1,1,1-Trichloroethane	µg/L		28.2	17.7 - 37.7	Not Reported	
5165	1,1,2-Trichloroethane	µg/L		0.00		Not Reported	
0057	Trichloroethylene	µg/L		63.2	40.2 - 82.5	Not Reported	
5175	Trichlorofluoromethane	µg/L		0.00		Not Reported	
5180	1,2,3-Trichloropropane (TCP)	µg/L		0.00		Not Reported	
5225	Vinyl acetate	µg/L		0.00		Not Reported	
5235	Vinyl chloride	µg/L		28.4	11.4 - 45.4	Not Reported	
5260	Xylenes, total	µg/L		150	86.0 - 201	Not Reported	

WP Low-Level PAHs (cat# 836)

5500	Acenaphthene	µg/L	7.20	8.61	4.70 - 9.94	Acceptable	EPA 8270C SIM
5505	Acenaphthylene	µg/L		5.40	2.61 - 6.45	Not Reported	
5555	Anthracene	µg/L		1.96	0.575 - 2.66	Not Reported	
5575	Benzo(a)anthracene	µg/L		1.42	0.885 - 1.59	Not Reported	
5585	Benzo(b)fluoranthene	µg/L		1.72	0.718 - 2.10	Not Reported	
5600	Benzo(k)fluoranthene	µg/L		0.394	0.158 - 0.534	Not Reported	
5590	Benzo(g,h,i)perylene	µg/L		0.939	0.333 - 1.24	Not Reported	
5580	Benzo(a)pyrene	µg/L		0.526	0.171 - 0.833	Not Reported	
5855	Chrysene	µg/L		0.834	0.469 - 1.08	Not Reported	
5895	Dibenz(a,h)anthracene	µg/L		0.979	0.278 - 1.40	Not Reported	
6265	Fluoranthene	µg/L	1.06	1.25	0.726 - 1.52	Acceptable	EPA 8270C SIM
6270	Fluorene	µg/L		4.03	1.67 - 4.86	Not Reported	
6315	Indeno(1,2,3-cd)pyrene	µg/L		1.43	0.656 - 1.78	Not Reported	
5005	Naphthalene	µg/L		9.64	3.91 - 10.6	Not Reported	
6615	Phenanthrene	µg/L		1.38	0.702 - 1.66	Not Reported	
6665	Pyrene	µg/L		0.442	0.226 - 0.592	Not Reported	





A Waters Company

WP-195 Final Complete Report

Carrie McCrillis
Quality Assurance Officer
Oilfield Environmental & Compliance
307 Roemer Way
Suite 300
Santa Maria, CA 93454
(805) 922-4772

EPA ID: CA01480
ERA Customer Number: O460853
Report Issued: 06/16/11
Study Dates: 04/11/11 - 05/26/11

Anal. No.	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description
-----------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------

WP Gasoline Range Organics (GRO) in Water (cat# 640)

9408	Gasoline Range Organics (GRO)	µg/L		2930	1140 - 5170	Not Reported	
4375	Benzene in GRO	µg/L		31.3	13.5 - 52.5	Not Reported	
4765	Ethylbenzene in GRO	µg/L	78.2	94.0	54.3 - 130	Acceptable	EPA 8021B
5140	Toluene in GRO	µg/L		273	147 - 364	Not Reported	
5260	Xylenes, total in GRO	µg/L	290	276	158 - 372	Acceptable	EPA 8021B

